

**COMMERCIAL - IN - CONFIDENCE**

# **CERTIFICATE OF CALIBRATION**

**on**  
**ONE ULTRASONIC FLOWMETER**  
of size 2" NB from  
**M/s. KRISTNAM TECHNOLOGY**  
**INDIAN INSTITUTE OF TECHNOLOGY**

**CERTIFICATE NUMBER**

**FCRI/WFL/C/2019/1226**  
**ULR-CC239519400001570F**

**एफ.सी.आर.आई**



**फ्लूइड कंट्रोल रिसर्च इंस्टिट्यूट, पालक्काड**  
**FLUID CONTROL RESEARCH INSTITUTE, PALAKKAD**

An Autonomous R&D Organisation under Ministry of Heavy Industries & Public Enterprises, Govt. of India.

**KANJIKODE WEST, PALAKKAD - 678 623, KERALA, INDIA.**

Ph : 91-391-2569010, 2566120, 2566206 Mob : 91-491-2566326 E-mail : [customercare@fcriindia.com](mailto:customercare@fcriindia.com) Web : [www.fcriindia.com](http://www.fcriindia.com)

COMMERCIAL - IN - CONFIDENCE

फ्लूइड कंट्रोल रिसर्च इंस्टिट्यूट, पालक्काड

FLUID CONTROL RESEARCH INSTITUTE, PALAKKAD

An Autonomous R&D Organisation under Ministry of Heavy Industries & Public Enterprises, Govt. of India.

KANJIKODE WEST, PALAKKAD - 678 623, KERALA, INDIA.

☎ 91-491-2569010, 2566120, 2566206 📠 91-491-2566326 ✉ customercare@fcriindia.com 🌐 www.fcriindia.com



**CERTIFICATE OF CALIBRATION**

OF ONE ULTRASONIC FLOWMETER  
of size 2" NB from  
M/s. KRISTNAM TECHNOLOGY  
INDIAN INSTITUTE OF TECHNOLOGY



Certificate No:  
CC-2395

Page 01 of 04 Pages

ULR-CC239519400001570F

*Water Flow Laboratory*

Date of Receipt
16.12.2019

Date of Calibration
31.12.2019

Date of issue
13.01.2020

CERTIFICATE NUMBER : FCRI/WFL/C/2019/1226

J C - 2019/2133

Approved Signatory

**SUMMARY**

**Test Meter**

ULTRASONIC FLOWMETER

**Standards Referred**

ISO 4185-1980: "Measurements of Liquid flow in closed conduits using weighing method"

**Laboratory**

Water Flow Laboratory (WFL)

**Calibration Results**

The results of calibration is given in Table 1, the calibration chart is given in Fig. 2 and summary of result in Section 3.

**Traceability**

All the instruments /Reference flow meters used are traceable to national standards through reference standards and their calibrations are valid.

NABL symbol on this certificate implies traceability of calibration data reported (Note 1, clause 5.6.2.1.1 of ISO 17025:2005).

Calibrated by

Prepared by

Checked by

K.G Jayesh

Amal Suresh

Fathima K A

K Suresh, SRE

ULR-CC239519400001570F

CERTIFICATE NUMBER : FCRI/WFL/C/2019/1226

PAGE 02 OF 04 PAGES

## 1. Method of Calibration

The meter to be calibrated was installed in a standard test line of water flow laboratory as shown in figure 1, Schematic of Calibration Setup. The line was flooded and entrapped air cleared using circuit air bleeds. Constant Head Tank / Direct pumping was the flow source for the calibration. Flow rate was adjusted using the downstream control valve. When flow conditions had stabilised, the flow rate was determined by collecting water for a measured time interval in the weighing system. The method used was flying-start-and-finish technique where the flow was diverted in to the weighing system and diverted back at the end of test.

The time of collection was determined by a high precision timer, which was triggered by a photo switch-timer blade arrangement attached to the diverter. The temperature of water was determined using on-line RTD's and density using an on-line Densitymeter. The actual flow rate and percentage error with respect to flowrate obtained from current output were then computed. This procedure was repeated for the other flow rates also.

## 2. Specification of Reference Instruments used

Instrument	Range	unit	Uncertainty	Calibration Due
Weighing System	2000	kg	1.00E-01 kg	30.12.2020
Timer	1000	sec	9.18E-04 sec	23.05.2020
Diverter System	1000	sec	6.28E-03 sec	04.04.2021
Densitymeter	3000	kg/m <sup>3</sup>	2.20E-02 kg/m <sup>3</sup>	06.09.2020
Data Acquisition Sysytem	4-20	mA	1.02E-02 mA	27.04.2020




Fig 1. Schematic of Calibration Set-up

Checked





COMMERCIAL - IN - CONFIDENCE

फ्लूइड कंट्रोल रिसर्च इंस्टिट्यूट, पालक्काड  
**FLUID CONTROL RESEARCH INSTITUTE, PALAKKAD**  
**CERTIFICATE OF CALIBRATION**



ULR-CC239519400001570F

CERTIFICATE NO  
 CC - 2395

CERTIFICATE NUMBER : FCRI/WFL/C/2019/1226

PAGE 03 OF 04 PAGES

Table. 1

FLOW ELEMENT : ULTRASONIC FLOWMETER  
 SIZE : 2" NB  
 MODEL NO : WFUBO2L0001  
 FLOW RANGE : 10 lps  
 CALIBRATING FLUID : WATER  
 DATE OF CALIBRATION : 31.12.2019

Sl. No.	Pup. bar	W <sub>1</sub> kg	W <sub>2</sub> kg	t sec	T deg C	ρ kg/m <sup>3</sup>	Q <sub>a</sub> m <sup>3</sup> /h	Q <sub>a</sub> lps	Q <sub>i</sub> lps	Error %	
1	2.00	1172.35	1684.45	98.896278	29.57	995.80	18.778	5.216	7.710	47.813	
2	2.00	290.95	868.40	57.894741	30.03	995.67	36.101	10.028	14.820	47.784	
3	2.00	112.70	1207.90	87.419055	30.17	995.63	45.347	12.596	18.780	49.089	
4	2.00	69.25	603.550	396.60505	30.80	995.47	4.877	1.355	2.06	52.058	
5	2.00	603.55	773.05	125.985126	30.85	995.46	4.871	1.353	2.06	52.258	
6	2.00	43.80	145.40	105.068676	31.00	995.43	3.501	0.972	0.968	-0.458	
7	2.00	145.40	247.05	106.634729	31.01	995.42	3.451	0.959	0.952	-0.694	
8	2.00	247.05	507.10	126.999552	31.02	995.42	7.413	2.059	2.065	0.279	
9	2.00	507.10	710.10	99.225856	31.03	995.42	7.407	2.057	2.077	0.951	
10	2.00	710.10	1231.90	177.051054	31.05	995.41	10.670	2.964	2.965	0.037	
11	2.00	1231.90	1728.40	168.499189	31.07	995.41	10.668	2.963	2.966	0.090	
12	2.00	175.20	603.05	107.102694	31.12	995.40	14.463	4.017	4.020	0.063	
13	2.00	603.05	1115.45	102.257102	31.14	995.40	18.142	5.039	4.995	-0.881	
14	2.00	1115.45	1608.65	82.148241	31.15	995.40	21.737	6.038	5.980	-0.959	
15	2.00	229.7	673.40	63.443989	31.15	995.40	25.323	7.034	6.935	-1.409	
16	2.00	673.40	1329.25	83.141811	31.16	995.40	28.559	7.933	7.870	-0.796	
17	2.00	1329.25	1953.10	69.483035	31.17	995.40	32.501	9.028	8.980	-0.532	
18	2.00	394.70	1076.40	68.741462	31.17	995.40	35.904	9.973	9.980	0.068	
19	2.00	713	1226.10	73.619781	31.18	995.40	25.253	7.015	6.97	-0.637	
Repeatability (%)									=	0.244	
Pup	- Pressure at the upstream of the test meter					Q <sub>i</sub>	- Indicated flowrate				
W <sub>1</sub>	- Initial mass of the weigh tank					Q <sub>a</sub> , Actual flowrate	= $\frac{(W_2 - W_1) \times 1.00106 \times 1000}{t \times \rho}$ lps				
W <sub>2</sub>	- Final mass of the weigh tank					% Dev in Q <sub>c</sub>	= $\frac{(Q_i - Q_a)}{Q_a} \times 100$ %				
B	- Buoyancy correction factor = 1.00106										
T	- Online temperature of water										
t	- Time of collection of water										
ρ	- Density of water at line temperature										

3. RESULT

- The results are tabulated in Table 1 and shown graphically in fig. 2.
- The expanded uncertainty in flowrate, taking into account the uncertainty of curve fit equation in Fig.2, is estimated to be better than **0.037 lps**
- The expanded uncertainty quoted are standard uncertainty multiplied by a coverage factor k = 2 at a level of confidence of approximately 95 %.

Checked



COMMERCIAL - IN - CONFIDENCE

फ्लूइड कंट्रोल रिसर्च इंस्टिट्यूट, पालक्काड  
FLUID CONTROL RESEARCH INSTITUTE, PALAKKAD  
CERTIFICATE OF CALIBRATION




ULR-CC239519400001570F

CERTIFICATE NO  
CC - 2395

CERTIFICATE NUMBER : FCRI/WFL/C/2019/1226

PAGE 04 OF 04 PAGES



*Checked*



**COMMERCIAL - IN - CONFIDENCE**

**फ्लुइड कंट्रोल रिसर्च इंस्टिट्यूट, पालक्काड  
FLUID CONTROL RESEARCH INSTITUTE, PALAKKAD  
CERTIFICATE OF CALIBRATION**



CERTIFICATE NO  
CC - 2395

एफ.सी.आर.आई

This page is intentionally kept

**BLANK**



**COMMERCIAL - IN - CONFIDENCE**

# **CERTIFICATE OF CALIBRATION**

**on**  
**ONE ULTRASONIC FLOWMETER**  
of size 4" NB from  
**M/s. KRISTNAM TECHNOLOGY**  
**INDIAN INSTITUTE OF TECHNOLOGY**

**CERTIFICATE NUMBER**

**FCRI/WFL/C/2019/1227**  
**ULR-CC239519400001571F**

**एफ.सी.आर.आई**



**फ्लूइड कंट्रोल रिसर्च इंस्टिट्यूट, पालक्काड**  
**FLUID CONTROL RESEARCH INSTITUTE, PALAKKAD**

An Autonomous R&D Organisation under Ministry of Heavy Industries & Public Enterprises, Govt. of India.

**KANJIKODE WEST, PALAKKAD - 678 623, KERALA, INDIA.**

Ph : 91-391-2569010, 2566120, 2566206 Mob : 91-491-2566326 E-mail : [customercare@fcriindia.com](mailto:customercare@fcriindia.com) Web : [www.fcriindia.com](http://www.fcriindia.com)

COMMERCIAL - IN - CONFIDENCE

फ्लूइड कंट्रोल रिसर्च इंस्टिट्यूट, पालक्काड

FLUID CONTROL RESEARCH INSTITUTE, PALAKKAD

An Autonomous R&D Organisation under Ministry of Heavy Industries & Public Enterprises, Govt. of India  
KANJIKODE WEST, PALAKKAD - 678 623, KERALA, INDIA.

☎ 91-491-2569010, 2566120, 2566206 📠 91-491-2566326 📧 customercare@fcriindia.com 🌐 www.fcriindia.com



**CERTIFICATE OF CALIBRATION**

OF ONE ULTRASONIC FLOWMETER  
of size 4" NB from  
M/s. KRISTNAM TECHNOLOGY  
INDIAN INSTITUTE OF TECHNOLOGY



Certificate No:  
CC-2395

Page 01 of 04 Pages

ULR-CC239519400001571F

*Water Flow Laboratory*

Date of Receipt
16.12.2019

Date of Calibration
24.12.2019

Date of issue
13.01.2020

CERTIFICATE NUMBER : FCRI/WFL/C/2019/1227

J C - 2019/2133

Approved Signatory

**SUMMARY**

**Test Meter**

ULTRASONIC FLOWMETER

**Standards  
Referred**

ISO 4185-1980: "Measurements of Liquid flow in closed conduits using weighing method"

**Laboratory**

Water Flow Laboratory (WFL)

**Calibration  
Results**

The results of calibration is given in Table 1, the calibration chart is given in Fig. 2 and summary of result in Section 3.

**Traceability**

All the instruments /Reference flow meters used are traceable to national standards through reference standards and their calibrations are valid.

NABL symbol on this certificate implies traceability of calibration data reported (Note 1, clause 5.6.2.1.1 of ISO 17025:2005).

Calibrated by

Prepared by

Checked by

K.G Jayesh

Amal Suresh

Fathima K A

K.Suresh, SRE





COMMERCIAL - IN - CONFIDENCE

फ्लूइड कंट्रोल रिसर्च इंस्टिट्यूट, पालक्काड  
**FLUID CONTROL RESEARCH INSTITUTE, PALAKKAD**  
**CERTIFICATE OF CALIBRATION**



CERTIFICATE NO  
 CC - 2395

ULR-CC239519400001571F

CERTIFICATE NUMBER : FCRI/WFL/C/2019/1227

PAGE 02 OF 04 PAGES

### 1. Method of Calibration

The meter to be calibrated was installed in a standard test line of water flow laboratory as shown in figure 1, Schematic of Calibration Setup. The line was flooded and entrapped air cleared using circuit air bleeds. Constant Head Tank / Direct pumping was the flow source for the calibration. Flow rate was adjusted using the downstream control valve. When flow conditions had stabilised, the flow rate was determined by collecting water for a measured time interval in the weighing system. The method used was flying-start-and-finish technique where the flow was diverted in to the weighing system and diverted back at the end of test.

The time of collection was determined by a high precision timer, which was triggered by a photo switch-timer blade arrangement attached to the diverter. The temperature of water was determined using on-line RTD's and density using an on-line Densitymeter. The actual flow rate and percentage error with respect to flowrate obtained from current output were then computed. This procedure was repeated for the other flow rates also.

### 2. Specification of Reference Instruments used

Instrument	Range	unit	Uncertainty	Calibration Due
Weighing System	20000	kg	2.90E+00 kg	02.12.2020
Timer	1000	sec	9.02E-03 sec	21.05.2020
Diverter System	1000	sec	8.10E-03 sec	11.08.2020
Densitymeter	3000	kg/m <sup>3</sup>	2.60E-02 kg/m <sup>3</sup>	06.09.2020
Data Acquisition System	4-20	mA	1.02E-02 mA	27.04.2020




Fig 1. Schematic of Calibration Set-up

*Checked*



COMMERCIAL - IN - CONFIDENCE

फ्लूइड कंट्रोल रिसर्च इंस्टिट्यूट, पालक्काड  
FLUID CONTROL RESEARCH INSTITUTE, PALAKKAD  
CERTIFICATE OF CALIBRATION



ULR-CC239519400001571F

CERTIFICATE NO  
CC - 2395

CERTIFICATE NUMBER : FCRI/WFL/C/2019/1227

PAGE 03 OF 04 PAGES

Table. 1

FLOW ELEMENT : ULTRASONIC FLOWMETER  
SIZE : 4" NB  
MODEL NO : WFUBO4L0001  
FLOW RANGE : 100 lps  
CALIBRATING FLUID : WATER  
DATE OF CALIBRATION : 24.12.2019

Sl No	Pup bar	W <sub>1</sub> kg	W <sub>2</sub> kg	t sec	T deg C	ρ kg/m <sup>3</sup>	Q <sub>a</sub> m <sup>3</sup> /h	Q <sub>a</sub> lps	Q <sub>i</sub> lps	Error %
1	2.00	6190	7492	143.575845	29.96	996.17	32.806	9.113	9.057	-0.613
2	2.00	7492	8468	107.797590	29.96	996.17	32.754	9.098	9.048	-0.554
3	2.00	8466	12354	63.052990	30.06	996.14	223.081	61.967	61.678	-0.466
4	2.00	10456	13678	50.943375	29.97	996.17	228.806	63.557	63.603	0.072
5	2.00	4420	10582	155.156959	30.14	995.84	143.722	39.923	39.692	-0.578
6	2.00	10582	16802	156.676221	30.14	995.84	143.668	39.908	39.659	-0.623
7	2.00	5118	11244	103.541969	30.19	995.82	214.113	59.476	59.277	-0.334
8	2.00	11244	17420	104.328027	30.19	995.82	214.234	59.509	59.331	-0.300
9	2.00	4464	10326	98.983806	30.23	995.77	214.331	59.536	59.653	0.196
10	2.00	10326	15442	86.444734	30.23	995.77	214.188	59.497	59.325	-0.289
11	2.00	11228	16880	95.477467	30.26	995.71	214.255	59.515	59.331	-0.310
12	2.00	4536	9848	100.489380	30.26	995.71	191.323	53.145	53.003	-0.268
13	2.00	9848	14844	94.532697	30.29	995.68	191.286	53.135	53.240	0.198
14	2.00	5580	10520	106.566138	30.29	995.68	167.784	46.607	46.711	0.224
15	2.00	10520	15446	106.349531	30.33	995.64	167.656	46.571	46.494	-0.166
16	2.00	5488	10222	143.222782	30.33	995.64	119.640	33.233	33.072	-0.485
17	2.00	10222	14780	137.999059	30.38	995.60	119.557	33.210	32.997	-0.642
18	2.00	14780	18322	134.285736	30.38	995.60	95.476	26.521	26.337	-0.695
19	2.00	10238	13886	138.596735	30.42	995.57	95.278	26.466	26.549	0.313
20	2.00	13886	16244	116.269953	30.42	995.57	73.412	20.392	20.254	-0.678
21	2.00	16244	18808	126.195264	30.46	995.53	73.550	20.431	20.475	0.217
22	2.00	10334	13434	231.109779	30.46	995.53	48.557	13.488	13.384	-0.771
23	2.00	13434	16880	256.326492	30.49	995.50	48.668	13.519	13.552	0.245

Linearity (%) = 0.542  
Repeatability (%) = 0.091

Pup - Pressure at the upstream of the test meter  
W<sub>1</sub> - Initial mass of the weigh tank  
W<sub>2</sub> - Final mass of the weigh tank  
B - Buoyancy correction factor = 1.00106  
T - Online temperature of water  
t - Time of collection of water  
ρ - Density of water at line temperature

Q<sub>i</sub> - Indicated flowrate

$$Q_a, \text{ Actual flowrate} = \frac{(W_2 - W_1) \times 1.00106 \times 1000}{t \times \rho} \text{ lps}$$

$$\% \text{ Dev in } Q_c = \frac{(Q_i - Q_a)}{Q_a} \times 100 \%$$

3. RESULT

- i. The results are tabulated in Table 1 and shown graphically in fig. 2
- ii. The expanded uncertainty in flowrate, taking into account the uncertainty of curve fit equation in Fig 2, is estimated to be better than 0.210 lps
- iii. The expanded uncertainty quoted are standard uncertainty multiplied by a coverage factor k = 2 at a level of confidence of approximately 95 %.

Checked



COMMERCIAL - IN - CONFIDENCE

फ्लूइड कंट्रोल रिसर्च इंस्टिट्यूट, पालक्काड  
FLUID CONTROL RESEARCH INSTITUTE, PALAKKAD  
CERTIFICATE OF CALIBRATION




ULR-CC239519400001571F

CERTIFICATE NO  
CC - 2395

CERTIFICATE NUMBER : FCRI/WFL/C/2019/1227

PAGE 04 OF 04 PAGES



Checked





**COMMERCIAL - IN - CONFIDENCE**

**फ्लूइड कंट्रोल रिसर्च इंस्टिट्यूट, पालक्काड**  
**FLUID CONTROL RESEARCH INSTITUTE, PALAKKAD**  
**CERTIFICATE OF CALIBRATION**




CERTIFICATE NO.  
CC - 2395

एफ.सी.आर.आई

This page is intentionally kept

**BLANK**



8/11/2024

**COMMERCIAL - IN - CONFIDENCE**

# **CERTIFICATE OF CALIBRATION**

**on**  
**ONE ULTRASONIC FLOW METER**  
**of size 6" NB from**  
**M/s. KRISTNAM TECHNOLOGY**  
**INDIAN INSTITUTE OF TECHNOLOGY**

**CERTIFICATE NUMBER**

**FCRI/WFL/C/2019/1228**  
**ULR-CC239519400001572F**

**एफ.सी.आर.आई**



**फ्लूइड कंट्रोल रिसर्च इंस्टिट्यूट, पालक्काड**

**FLUID CONTROL RESEARCH INSTITUTE, PALAKKAD**

An Autonomous R&D Organisation under Ministry of Heavy Industries & Public Enterprises, Govt. of India.

**KANJIKODE WEST, PALAKKAD - 678 623, KERALA, INDIA.**

Ph : 91-391-2569010, 2566120, 2566206 Mob : 91-491-2566326 E-mail : [customercare@fcriindia.com](mailto:customercare@fcriindia.com) Web : [www.fcriindia.com](http://www.fcriindia.com)

COMMERCIAL - IN - CONFIDENCE

फ्लूइड कंट्रोल रिसर्च इंस्टिट्यूट, पालक्काड

FLUID CONTROL RESEARCH INSTITUTE, PALAKKAD

An Autonomous R&D Organisation under Ministry of Heavy Industries & Public Enterprises, Govt. of India.

KANJIKODE WEST, PALAKKAD - 678 623, KERALA, INDIA.

☎ 91-491-2569010, 2566120, 2566206 📠 91-491-2566326 📧 customercare@fcriindia.com 🌐 www.fcriindia.com



## CERTIFICATE OF CALIBRATION

OF ONE ULTRASONIC FLOW METER  
of size 6" NB from  
M/s. KRISTNAM TECHNOLOGY  
INDIAN INSTITUTE OF TECHNOLOGY



Certificate No:  
CC-2395

Page 01 of 04 Pages

ULR-CC239519400001572F

*Water Flow Laboratory*

Date of Receipt
16.12.2019

Date of Calibration
13.12.2019

Date of issue
13.01.2020

CERTIFICATE NUMBER : FCRI/WFL/C/2019/1228

J C - 2019/2133

Approved Signatory

### SUMMARY

**Test Meter**

ULTRASONIC FLOW METER

**Standards Referred**

ISO 4185-1980: "Measurements of Liquid flow in closed conduits using weighing method"

**Laboratory**

Water Flow Laboratory (WFL)

**Calibration Results**

The results of calibration is given in Table 1, the calibration chart is given in Fig. 2 and summary of result in Section 3.

**Traceability**

All the instruments /Reference flow meters used are traceable to national standards through reference standards and their calibrations are valid.

NABL symbol on this certificate implies traceability of calibration data reported (Note 1, clause 5.6.2.1.1 of ISO 17025:2005).

Calibrated by

Prepared by

Checked by

K.G Jayesh

Fathima K A

K Suresh, SRE

Amal Suresh





**COMMERCIAL - IN - CONFIDENCE**  
**फ्लूइड कंट्रोल रिसर्च इंस्टिट्यूट, पालक्काड**  
**FLUID CONTROL RESEARCH INSTITUTE, PALAKKAD**  
**CERTIFICATE OF CALIBRATION**



CERTIFICATE NO  
CC - 2395

ULR-CC239519400001572F

CERTIFICATE NUMBER : FCRI/WFL/C/2019/1228

PAGE 02 OF 04 PAGES

## 1. Method of Calibration

The meter to be calibrated was installed in a standard test line of water flow laboratory as shown in figure 1, Schematic of Calibration Setup. The line was flooded and entrapped air cleared using circuit air bleeds, Constant Head Tank / Direct pumping was the flow source for the calibration. Flow rate was adjusted using the downstream control valve. When flow conditions had stabilised, the flow rate was determined by collecting water for a measured time interval in the weighing system. The method used was flying-start-and-finish technique where the flow was diverted in to the weighing system and diverted back at the end of test.

The time of collection was determined by a high precision timer, which was triggered by a photo switch-timer blade arrangement attached to the diverter. The temperature of water was determined using on-line RTD's and density using an on-line Densitometer. The actual flow rate and percentage error with respect to flowrate obtained from current output were then computed. This procedure was repeated for the other flow rates also.

## 2. Specification of Reference Instruments used

Instrument	Range	unit	Uncertainty	Calibration Due
Weighing System	20000	kg	2.90E+00 kg	02.12.2020
Timer	1000	sec	9.02E-03 sec	21.05.2020
Diverter System	1000	sec	8.10E-03 sec	11.08.2020
Densitometer	3000	kg/m <sup>3</sup>	2.60E-02 kg/m <sup>3</sup>	06.09.2020
Data Acquisition Sysytem	4-20	mA	1.02E-02 mA	27.04.2020




Fig 1, Schematic of Calibration Set-up

*Checked*



COMMERCIAL - IN - CONFIDENCE

फ्लूइड कंट्रोल रिसर्च इंस्टिट्यूट, पालक्काड  
FLUID CONTROL RESEARCH INSTITUTE, PALAKKAD  
CERTIFICATE OF CALIBRATION



ULR-CC239519400001572F

CERTIFICATE NO  
CC - 2395

CERTIFICATE NUMBER : FCRI/WFL/C/2019/1228

PAGE 03 OF 04 PAGES

Table. 1

FLOW ELEMENT : ULTRASONIC FLOW METER  
SIZE : 6" NB  
MODEL NO : WFUBO6L0001  
FLOW RANGE : 100 lps  
CALIBRATING FLUID : WATER  
DATE OF CALIBRATION : 13.12.2019

Sl. No	Pup bar	W <sub>1</sub> kg	W <sub>2</sub> kg	t sec	T deg C	ρ kg/m <sup>3</sup>	Q <sub>a</sub> m <sup>3</sup> /h	Q <sub>a</sub> lps	Q <sub>i</sub> lps	Error %
1	2.00	1368	10000	96.104514	29.58	996.56	324.808	90.224	90.79	0.627
2	2.00	10000	17444	82.929516	29.58	996.56	324.606	90.168	90.74	0.634
3	2.00	2982	9568	81.965097	29.59	996.55	290.574	80.715	81.20	0.603
4	2.00	9566	16134	81.738767	29.59	996.55	290.582	80.717	81.22	0.623
5	2.00	1910	8234	90.130024	29.62	996.54	253.741	70.484	70.62	0.187
6	2.00	8232	14512	89.580094	29.65	996.54	253.522	70.423	70.87	0.635
7	2.00	2084	7166	85.824647	29.67	996.53	214.139	59.483	59.86	0.632
8	2.00	7164	13534	107.643292	29.68	996.53	214.005	59.446	59.84	0.668
9	2.00	9548	15976	162.126585	29.70	996.52	143.383	39.829	40.03	0.499
10	2.00	1824	7198	135.511713	29.71	996.52	143.416	39.838	40.05	0.539
11	2.00	7196	11238	417.985891	29.76	996.51	34.972	9.714	9.70	-0.158
12	2.00	11236	15140	404.296772	29.79	996.51	34.921	9.700	9.65	-0.519
13	2.00	2538	9156	67.116319	29.82	996.49	356.606	99.057	99.765	0.714
14	2.00	9152	19006	99.990047	29.84	996.49	356.406	99.0018	99.75	0.756
15	2.00	5080	11030	118.073711	29.34	996.62	182.220	50.617	50.850	0.461
16	2.00	11028	16664	111.849814	29.35	996.62	182.209	50.613	50.890	0.546
17	2.00	2476	8690	122.787070	29.85	996.48	183.026	50.8405	51.18	0.658
18	2.00	8690	15112	126.969125	29.85	996.48	182.922	50.8117	51.07	0.508
19	2.00	2058	8340	206.367380	29.86	996.48	110.091	30.5808	30.75	0.553
20	2.00	8338	14292	195.756822	29.87	996.48	109.998	30.5551	30.71	0.507
21	2.00	3644	7064	171.311927	29.88	996.48	72.199	20.0553	20.15	0.472
22	2.00	7064	11062	200.176605	29.89	996.47	72.232	20.0644	20.09	0.128
23	2.00	4042	10430	128.718892	29.90	996.47	179.482	49.8561	50.01	0.317

Linearity (%) = 0.637  
Repeatability (%) = 0.042

Pup - Pressure at the upstream of the test meter  
W<sub>1</sub> - Initial mass of the weigh tank  
W<sub>2</sub> - Final mass of the weigh tank  
B - Buoyancy correction factor = 1.00106  
T - Online temperature of water  
t - Time of collection of water  
ρ - Density of water at line temperature

Q<sub>i</sub> - Indicated flowrate

$$Q_a, \text{ Actual flowrate} = \frac{(W_2 - W_1) \times 1.00106 \times 1000}{t \times \rho} \text{ lps}$$

$$\% \text{ Dev in } Q_c = \frac{(Q_i - Q_a)}{Q_a} \times 100$$

3. RESULT

- The results are tabulated in Table 1 and shown graphically in fig. 2
- The expanded uncertainty in flowrate, taking into account the uncertainty of curve fit equation in Fig.2, is estimated to be better than **0.119 lps**
- The expanded uncertainty quoted are standard uncertainty multiplied by a coverage factor k = 2 at a level of confidence of approximately 95 %.

Checked



COMMERCIAL - IN - CONFIDENCE

फ्लूइड कंट्रोल रिसर्च इंस्टिट्यूट, पालक्काड  
FLUID CONTROL RESEARCH INSTITUTE, PALAKKAD  
**CERTIFICATE OF CALIBRATION**




ULR-CC239519400001572F

CERTIFICATE NO  
CC - 2395

CERTIFICATE NUMBER : FCRI/WFL/C/2019/1228

PAGE 04 OF 04 PAGES



*Checked*





**COMMERCIAL - IN - CONFIDENCE**

**फ्लुइड कंट्रोल रिसर्च इंस्टिट्यूट, पालक्काड  
FLUID CONTROL RESEARCH INSTITUTE, PALAKKAD  
CERTIFICATE OF CALIBRATION**



CERTIFICATE NO  
CC - 2395

एफ. सी. आर. आई

This page is intentionally kept

**BLANK**